



भाकृअनुप-राष्ट्रीय उच्च सुरक्षा पशुरोग संस्थान  
ICAR-National Institute of High Security Animal Diseases  
(एवियन इन्फ्लूएंजा ओ.आई.ई. संदर्भ प्रयोगशाला)  
(OIE Reference Laboratory for Avian Influenza)  
आनंद नगर, भोपाल - ४६२०२२ (म.प्र.), भारत  
Anand Nagar, Bhopal - 462022 (M.P.), India



EPABX Tel. No. 0755-2757542, 2750647, FAX: 0755-2758842, Website: www.nihsad.nic.in

F. No. 9-322/16-17/NIHSAD (P&S)

Dated: - 13.02.2017

To

.....  
.....

Sub:- Purchase of Egg Incubator under CRP Project – reg.

Dear Sirs,

We intend to purchase Egg Incubator with the following specifications: -

Description	Qty required
Egg Incubator Technical Specification Attached Annexure – I	01 No.

You are requested to please give **Quotation** of the above mentioned item. Please note that: -

1. Rates are to be quoted on F.O.R. NIHSAD, Bhopal basis.
2. Warranty for supply is to be offered for reasonable period (at least one year). The firm ordered for equipment will have to deposit 10% Security for the period of Warranty before supply.
3. Taxes, if any, to be shown separately.
4. E-Payment will be made within 30 days from the date of receiving of goods in satisfactory condition & it's Installation.
5. Date & Time of Opening (Tentative):- **02.03.2017 at 11:30 AM.**
6. Bid Validity Period: - 120 days.
7. Delivery Period: 10 days.

The Quotation may be submitted upto **01<sup>st</sup> March, 2017** in a closed envelope duly superscribed.

Yours Sincerely,  
*Bramhal*  
13/02/2017  
Sr. Adm. Officer

**Annexure – I**

**Technical Specification for Egg Incubator**

**Egg Incubator  
Specification-**

- ✓ Incubator for poultry eggs,
- ✓ Compact unit with observation window,
- ✓ Capacity upto 250 eggs,
- ✓ Fully automated for control of temp (30-40°C or wider range) and humidity (80-90% relative humidity),
- ✓ Digital display,
- ✓ Temperature measurement accuracy:  $\pm 0.1^{\circ}\text{C}$ .,
- ✓ Humidity control accuracy:  $\pm 5\% \text{RH}$ ..
- ✓ Automatic Egg turning,
- ✓ Equipment shall be with all safety features as per BIS,
- ✓ Comprehensive warranty of at least 2 years (onsite),
- ✓ Quote should include the following spares:
  1. Temperature and humidity controller -1 Nos.
  2. Solenoid valve -1Nos,
  3. Vee Belt- I Nos.