



भाकृअनुप-राष्ट्रीय उच्च सुरक्षा पशुरोग संस्थान  
ICAR-National Institute of High Security Animal Diseases

(एवियन इन्फ्लूएंजा ओ.आई.ई. संदर्भ प्रयोगशाला)  
(OIE Reference Laboratory for Avian Influenza)

आनंद नगर, भोपाल - ४६२०२२ (म.प्र.), भारत  
Anand Nagar, Bhopal - 462022 (M.P.), India



EPABX Tel. No. 0755-2757542, 2750647, FAX: 0755-2758842, Website: www.nihsad.nic.in

Dated: - 29.05.2017

F. No. 9-28/17-18/NIHSAD (P&S)

To

.....  
.....

Sub:- Purchase of Pressure Recoder for ETP – reg.

Dear Sirs,

We intend to purchase Pressure Recoder for ETP with the following specifications: -


Description	Qty required
Pressure Recoder Technical Specification Attached Annexure – I	01 No.

You are requested to please give **Quotation** of the above mentioned item. Please note that: -

1. Rates are to be quoted on F.O.R. NIHSAD, Bhopal basis.
2. Warranty for supply is to be offered for reasonable period (at least one year). The firm ordered for equipment will have to deposit 10% Security for the period of Warranty before supply.
3. Taxes, if any, to be shown separately.
4. E-Payment will be made within 30 days from the date of receiving of goods in satisfactory condition & it's Installation.
5. Date & Time of Opening (Tentative):- **20.06.2017 at 11:30 AM.**
6. Bid Validity Period: - 120 days.
7. Delivery Period: 10 days.

The Quotation may be submitted upto **19<sup>th</sup> June, 2017** in a closed envelope duly superscribed.

Yours Sincerely,

  
(S. K. GUPTA)  
Sr. Adm. Officer

**Specification for Pressure Recorder**

<b>Make</b>	<b>Yokogawa or Honeywell</b>
<b>Type:</b>	Pressure Recorder (paperless)
<b>Display:</b>	5.7" Digital Color LCD (TFT) Touch screen
<b>Panel cutout size</b>	138MM(H) X 138MM
<b>Dimensions: Bezel.</b>	W: 144mm (5.67"), H: 144mm (5.67"), D: 200/280mm , (7.87")
<b>Analog Input</b>	06 Analog Inputs (minimum), 4-20 mA DC
<b>Expansion card:</b>	Optional
<b>Power:</b>	100 - 240 VAC 50/60 Hz
<b>Transmitter power option:</b>	<b>24Vdc Transmitter Power,</b>
<b>Memory:</b>	256 MB(minimum)
<b>Mounting:</b>	Standard Panel Mounting
<b>USB interface</b>	2 nos. To connect keyboard, mouse, External HDD,
<b>Operating Temperature</b>	0 to 50 Deg C
<b>Buzzer Sound</b>	Required in instrument
<b>Accuracy:</b>	±0.1% of FS (Included Tolerance)
<b>Communications</b>	Through Ethernet
<b>Warranty On Site</b>	2 years from date of installation & commissioning
<b>Note</b>	<p>1) Bidders are advised in their own interest to visit site as input is coming from Honeywell pressure transmitter to Yokogawa make chart recorder model uR1000.</p> <p>2) Payment will release after successfully installation and displaying 6 channel pressure data on recorder by supplier as we are replacing Yokogawa make chart recorder model uR1000 by chartless recorder</p>

*Deep Kumar Gupta*