



भाकृअनुप-राष्ट्रीय उच्च सुरक्षा पशुरोग संस्थान
ICAR-National Institute of High Security Animal Diseases

(एवियन इन्फ्लूएंजा ओ.आई.ई. संदर्भ प्रयोगशाला)

(OIE Reference Laboratory for Avian Influenza)

आनंद नगर, भोपाल - ४६२०२२ (म.प्र.), भारत

Anand Nagar, Bhopal - 462022 (M.P.), India



EPABX Tel. No. 0755-2757542, 2750647, FAX: 0755-2758842, Website: www.nihsad.nic.in

F. No. 9-173/17-18/NIHSAD (P&S)

Dated: 07.12.2017

To,

Sub:- Quotation (**Under two bid system**) for purchase of Hot Air Oven – reg.

Dear Sir(s),

Reputed firms/Authorized firms are invited for quoting their rates for the supply of following stores as per under mentioned specifications:-

Name of item	Qty.	Rate (In Rupees)	Total (In Rupees)
Hot Air Oven Specification attached	01		
GST@ 18%			
*Total FOR NIHSAD, Bhopal			

***(Financial bid must be submitted in separate envelope duly sealed)**

Relevant terms & conditions are as below: -

01. Quotations should be submitted **under two bid system** in a separate envelope (**Technical Bid + Financial Bid**) clearly mentioning at the top Tender No. F. No 9-173/17-18/NIHSAD (P&S) "Hot Air Oven" and EMD particulars (if any). The Technical bid and the Financial bid should be sealed in a separate envelopes and both these sealed covers are to be put in a bigger cover which should also be sealed. The **last date** of receipt of above quotation is **28.12.2017 upto 5:00 PM** This institute will not be responsible for any postal delays.
02. The rates should be quoted for F.O.R. ICAR-NIHSAD, Bhopal basis along with indicating separately brand name its price, packing forwarding, freight, tax percentage and insurance etc. in clear terms.
03. Discount, if any, should be indicated clearly.
04. This institute is registered with the Department of Scientific & Industrial Research (DSIR), Govt. of India for purpose of availing custom duty exemption in terms of Government Notification No. 51/96-Customs dated 23 July 1996 and central excise duty exemption in terms of Government Notification No. 10/97-Central Excise dated 1 March 1997.
05. GST should be indicated in clear terms separately. GSTIN Registration/TIN Number should be mentioned.
06. Unless otherwise mentioned in the quotation, it will be presumed that the quoted rates are inclusive of all taxes/levies and free delivery/installation at our Institute.

Contd....2

07. The rates quoted should be valid for **180 days** from the date of opening of quotation.
08. There should be no cutting/ overwriting. The cutting if any should be duly attested.
09. Warranty – 02 Years from the date of Installation.
10. Quotation should be submitted along with printed literature/ catalogue of the quoted make/ model and authorization certificate of the manufacturer.
11. Unattested/amended/overwritten figures will not be considered.
12. In all matters of disputes, the decision of the Director of this Institute shall be final and binding.
13. Minimum period of delivery (not more than 20 days) should clearly be specified in the quotation.
14. TDS will be deducted from the bill as per applicable tax rules.
15. PAYMENT TERMS: Payment will be made within reasonable time say within 30 days from the date of receipt of material in good condition without any damage and after installation of the equipment in accordance with the ordered specification at the end. **Since the mode of payment will be e-payment system, hence Bank Account Number, name of bank with location and IFSC Code Number of Bank may be furnished on the body of the bill.**
16. **Bid Security (EMD):** Rs. 5000/- must be submitted along with the quotation in the form of Demand Draft/FDR/Banker's Cheque drawn **in favour of "ICAR UNIT, NIHSAD, Bhopal"** otherwise the bid will not be taken into consideration. Micro & Small Enterprises as defined in MSE Procurement Policy are exempted for producing the EMD.
17. Performance Security: Within 20 (Twenty) days after the issue of a letter/PO by this Institute, the supplier, shall furnish performance security to the institute for an amount of **10% (of Total Ordered value)** including the warranty obligation before supply.
18. The Institute reserves the right to reject any or all the quotation (s) received without assigning any reason thereof.
19. Tentative date of opening of Technical bid 29.12.2017 at 3:30 PM.

Yours sincerely,


Adm. Officer

REVISED SPECIFICATION FOR HOT AIR OVEN

1	Material	Double walled construction with outer of MS with powder coated and inner made 304 quality stainless steel sheet.
2	Chamber Volume	Minimum of 150 lit
3	Insulation	Thick Ceramic wool/ Glass wool insulation on all sides including door.
4	Door	Fitted with hinges and door lock.
5	Gasket	Gasket made of silicone rubber/ viton for leak-proof operation.
6	Trays	Minimum two, Adjustable rod mesh or perforated trays, stainless steel.
7	Heathers	Stainless steel Nichrome wire air heathers.
8	Air Circulation	Forced Air circulating System with fan for temperature Uniformity.
9	Ventilator	Air Ventilator.
10	Sensor type	RTD / PT-100 sensor.
11	Controller	Microprocessor based digital PID controller with both process and set value display.
12	Operation	220±10 V, single phase, Ac supply
13	Temperature range	5°C above ambient to 250°C.
14	Accuracy	+/- 1 °C
15	Safety	Temperature safety as per DIN 12880 Class 3.1
16	Timer	Digital presettable with range 0-999 min.
17	Alarm	Visual and acoustic alarm if temperature exceeds the set points.
18	Certification	CE/ EQUIVALENT certification for the quoted model.
19	Warranty	2 years
20	User list	Should be provided with a list of users in India (particularly Govt. of India R & D Organizations) and two performance certificate from institutes preferably government institute for similar instrument supplied 2 year back towards its performance & services should be included with the offer.