



भाकृअनुप-राष्ट्रीय उच्च सुरक्षा पशुरोग संस्थान
ICAR-National Institute of High Security Animal Diseases

(एवियन इन्फ्लूएंजा ओ.आई.ई. संदर्भ प्रयोगशाला)
(OIE Reference Laboratory for Avian Influenza)
आनंद नगर, भोपाल - ४६२०२२ (म.प्र.), भारत
Anand Nagar, Bhopal - 462022 (M.P.), India



EPABX Tel. No. 0755-2757542, 2750647, FAX: 0755-2758842, Website: www.nihsad.nic.in

F. No. 9-339/16-17/NIHSAD (P&S)

Dated: - 13.02.2017

To

.....
.....

Sub:- Purchase of Quartz Water Double Distillation Unit under CRP Project – reg.

Dear Sirs,

We intend to purchase Quartz Water Double Distillation Unit with the following specifications: -

Description	Qty required
Quartz Water Double Distillation Unit Technical Specification Attached Annexure – I	01 No.

You are requested to please give **Quotation** of the above mentioned item. Please note that: -

1. Rates are to be quoted on F.O.R. NIHSAD, Bhopal basis.
2. Warranty for supply is to be offered for reasonable period (at least one year). The firm ordered for equipment will have to deposit 10% Security for the period of Warranty before supply.
3. Taxes, if any, to be shown separately.
4. E-Payment will be made within 30 days from the date of receiving of goods in satisfactory condition & it's Installation.
5. Date & Time of Opening (Tentative):- **02.03.2017 at 11:30 AM.**
6. Bid Validity Period: - 120 days.
7. Delivery Period: 10 days.

The Quotation may be submitted upto **01st March, 2017** in a closed envelope duly superscribed.

Yours Sincerely
R. S. Mishra
13/02/2017
Sr. Adm. Officer / AAD

F. No. 9-339/16-17/NIHSAD (P&S)
Dated: - 13.02.2017

Annexure – I

Technical Specifications for Quartz water distillation

1. Double stage of Distillation.
2. Output water should have Conductivity $< 1 \mu\text{S}$ and pH should be in the range of pH: 6.0-7.0.
3. Quartz distiller, Demountable boiler for easy cleaning and maintenance.
4. Minimum 3-4 liters per hour output.
5. Self-standing with all the accessory needed for.
6. Electrical dual auto cut-off system for in case water level increases or decrease and also if Boiler pressure increases, Water Flow switch-Incase water level reduces or fails through Condenser, Alarm Audio for any breakdown.
7. Additional Bulb Condenser to bring down the temperature of the distillate.
8. MCB and all electrical items should be BIS approved.
9. Power Supply: $230 \text{ V} \pm 10\%$, 50Hz.
10. Any accessory needed for installation and commissioning of the system should be included in the offer.