



F. No. 9-89/21-22/NIHSAD (P&S)

Dated: - 06.12.2021

CORRIGENDUM

Sub:- Pre Bid Meeting for the Purchase of Bio-containment individually ventilated cage system with fittings at ICAR-NIHSAD, Bhopal – reg.

Ref:- Tender ID: - 2021_DARE_659080_1

The pre bid conference with regard to purchase of Bio-Containment individually ventilated cage system with fittings at ICAR-NIHSAD, Bhopal was held on 02.12.2021 in online mode as well as physical mode. As per the Queries raised/discussion in the conference, the following points are clarified. The bidders may take into account these points also and submit the bids accordingly:-

	Original Specification	Revised Specification
1.	Cages and Bottles: 6.Cage should be fitted with Microbiological filter at cage level, to ensures optimal gas exchange as well as protection during procedures or in case of power failure.	Cages and Bottles: 6. Cage should be fitted with HEPA filters with efficiency of at least 99.97% at cage level to ensure maximum safety for the personnel during procedures or in case of power failure.
2.	2. The cage top should be fitted with appropriate and efficient gasket sealing to prevent leakage of air. Cages should be tested for leakages.	2. The cages including bottle fittings should be hermetically sealed appropriately to prevent leakage of air. Cages should be demonstrated to be leak proof during onsite installation.
3.	Other terms and conditions 6. 5 years Comprehensive Maintenance cost beyond warranty period also to be quoted for the Bio-containment IVC system	Other terms and conditions 6. Annual Maintenance cost for five years beyond the warranty period also to be quoted for the Bio-containment IVC system


(PS Sunil Kumar)

Admn. Officer

Phone No. 0755-2754672