



भा.कृ.अ.प.-राष्ट्रीय उच्च सुरक्षा पशुरोग संस्थान
ICAR-National Institute of High Security Animal Diseases

(एवियन इन्फ्लूएंजा ओ.आई.ई. संदर्भ प्रयोगशाला)

(OIE Reference Laboratory for Avian Influenza)

आनंद नगर, भोपाल - ४६२०२२ (म.प्र.), भारत

Anand Nagar, Bhopal - 462022 (M.P.), India



EPABX Tel. No. 0755-2757542, 2750647, FAX: 0755-2758842, Website: www.nihsad.nic.in

F. No. 9-16/21-22/NIHSAD (P&S)

Dated: - 17.11.2021

NOTICE
(For Technical Suitability)


With reference to Tender ID 2021_DARE_642168_1 for the purchase of Deep Freezer (-80°C), the following three firms have been found to be technically suitable and the financial bids will be opened on 22.11.2021 at 11:30 AM.

Sr. No.	Name of firm	Quoted Make	Quoted Model
1.	M/s. Cryo Scientific Systems Pvt. Ltd., Chennai	Cryo Scientific	URC-V-650-4
2.	M/s. Chirag Scientific Sales, Ganj Basoda	Eppendorf	Cryo Cube F570h
3.	M/s. Helix Biosciences, New Delhi	Eppendorf	CryoCube F740h

The following sixteen firms are not technically suitable as per the details given below: -

S. No	Bidder	Reason for rejection
1	M/s. Adarsh Enterprises, Jabalpur	Only 2 performance certificate submitted
2	M/s. Bioaid Lab Solutions, Bhopal	No vacuum insulation panel in the quoted model. Only PUF
3	M/s. Biolinkk, New Delhi	Only one compressor in the model. No performance certificate submitted.
4	M/s. Biotechnologies Inc, New Delhi	No performance certificate submitted.
5	M/s. Delta Technologies, Kolkata	No vacuum insulation panel in the quoted model. No performance certificate submitted.
6	M/s. Jindal Enterprises, Bhopal	No vacuum insulation panel in the quoted model. No performance certificate submitted.
7	M/s. Jogeshwari Engg Industries Nashik	Model not clearly specified. No Performance certificate not attached
8	M/s. Labtronix Inc, Faridabad	No vacuum insulation panel in the quoted model. No performance certificate submitted.
9	M/s. Reckers Automation India, Delhi	No performance certificate submitted.
10	M/s. Riransa Inc Healthcare, Lucknow	No vacuum insulation panel in the quoted model. No performance certificate submitted.
11	M/s. S.M. Scientific Instruments, Delhi	Details of volume and compartments not clearly specified
12	M/s. Saksham Technologies Pvt. Ltd., Delhi	No performance certificate submitted.
13	M/s. San Engg Company Aurangabad	No performance certificate submitted.
14	M/s. Scientific Systems & Chemicals, Bhopal	Only one performance certificate submitted.
15	M/s. Sun Syndicate, Coimbatore	Quoted model has less capacity (385 L against the required 500L or more) and fewer compartments (2 against the required 3 or more). No performance certificate submitted.
16	M/s. Sylis Technologies, Navi Mumbai	No performance certificate submitted.

It is for information to concerned bidders/firms



(PS Sunil Kumar)
Administrative Officer
Phone No. 0755-2754672